

Member State
P.R.China

OIML Certificate No.
R60/2000-CN-03.02

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: No.9 Madiandonglu, Haidian District, Beijing, China

Applicant

Name: Yuyao TongDa Scales Co., LTD.
Address: 21 South Heyan Road, Yuyao, Zhejiang Province, P.R. China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:
Address:

Identification of the certified type

Model designation	TD132								
Accuracy class	C3								
Maximum capacity (E_{max})	1.1t	1.5t	2t	3t	5t	2.5klb	4klb	5klb	10klb
Minimum dead load (E_{min})	0.1t	0.15t	0.2t	0.3t	0.5t	0.25klb	0.4klb	0.5klb	1klb
Maximum number of load cell intervals (n_{max})	3000								
Minimum load cell verification interval (V_{min})	160g	240g	320g	480g	800g	0.4lb	0.64lb	0.8lb	1.61lb
Apportionment factor (P_{LC})	0.7								
Ratio of minimum LC verification interval ($Y = E_{max}/V_{min}$)	6875	6250							
Safe overload	1.5t	2.25t	3t	4.5t	7.5t	3.75klb	6klb	7.5klb	15klb
Operating temperature range	-10°C - +40°C								
Humidity condition	CH								

The numbers of the associated test reports and the included pages:

LXff2003-6001 18pages



OIML CERTIFICATE OF CONFORMITY

To YUYAO TongDa Scales Co., LTD.

This certificate attests the conformity of the type of the Load
Cell TD132 with the requirements
of the Recommendation R60(2000) of the International
Organization of Legal Metrology(OIML).

The issuing authority

Office of OIML Affairs

中国秘书处

April 7, 2003

The OIML member

王秦平

April 7, 2003

This certificate attests the conformity of the above-mentioned type (represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C3

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports.

The issuing authority

Office of OIML Affairs

中国秘书处

April 7, 2003

The OIML member

王秦平

April 7, 2003

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.